

DIRECTIONS FOR REMOVAL:

Within 24 hours, soak GlykoSL[®] with water and gently remove.


Store at controlled room temperature,
15°C to 30°C (59°F to 86°F).

Sterilized with radiation.

Do not resterilize or reuse.

Single use only.

Not Made with Natural Rubber Latex



Marine Polymer
Technologies, Inc.
107 Water Street
Danvers, MA 01923
1-888-666-2560

GlykoSL[®]

REF 400-13

Distributed by:
Marine Polymer Technologies, Inc.
107 Water Street • Danvers, MA 01923
1-888-666-2560
Made in the U.S.A.

MPS 500-28 rev 04

CAUTION: Federal (USA) law restricts this device to sale by or on the order of a physician or other licensed practitioner.

DESCRIPTION:

GlykoSL[®] is a conformable, soft and absorbent calcium alginate dressing attached to a foam backing.

INDICATIONS:

GlykoSL[®] is indicated for use in the local management of bleeding wounds such as vascular access sites and percutaneous catheters or tubes.

PRECAUTIONS:

There are no known contraindications to the use of GlykoSL[®].

Longer compression may be necessary for hypertensive or obese patients.

DIRECTIONS FOR APPLICATION:

1. Open GlykoSL[®] pouch into sterile field.
2. Hold pressure proximal to the puncture site.
3. Place GlykoSL[®] over puncture site, **smooth side away from skin.**
4. Apply firm compression over the puncture site and release the proximal pressure after 3 to 4 minutes.
5. Hold pressure on puncture site until hemostasis is achieved, then slowly release.
6. Place dry gauze over GlykoSL[®] and cover with a transparent dressing.