

A LOOK AT YOUR BREEZE² METER:



Button Door Open



Back of Meter



1. Open Meter

Turn the meter over (display screen down) and open by pressing in on the back edge of the **open latch**, then pulling up.




2. Insert Disc

Insert a new 10-test disc — bumpy side up.




Close the meter and snap it shut.

IMPORTANT: Keep the meter flat.




3. Prepare Lancing Device

Remove the end-cap from the MICROLET² device.



Rotate (but do not remove) the protective cap on a lancet 1/4 turn.



Then insert the lancet firmly to cock the device.



Twist off the lancet cap and replace the endcap.




See your MICROLET² insert for complete instructions.

4. Push Out Test Strip

Pull and Push the meter handle to expose a test strip and automatically turn on the meter.




Note: The **release button** is **UP**, a test strip is pushed out and a series of “apply blood” prompts appears on the screen. The meter is now ready to run a blood glucose test.



5. Test

Stick your finger and form a small, **round** drop of blood.



Move the entire front edge of the test strip until it touches the blood drop. The blood is automatically pulled into the test strip and fills the strip completely.



Hold the test strip in place until the meter beeps and the countdown begins, then move it away.



The test result is displayed on the screen after the 5-second count-down and will be stored automatically in the meter for future use.



Hold the meter, pointing down, over a waste container and press the **release button**.




The used test strip will fall into the waste container.

Press the **ⓘ** button to turn the meter off and conserve battery power. (See *inside front cover for location*.)


TROUBLESHOOTING AND ERRORS:



There was a problem completely filling the test strip. Repeat the test with a new strip.



Your test result is above 600 mg/dL. Wash your hands and the test site, then repeat the test. If **HI** appears again, contact your physician or healthcare professional.



Battery is low. Replace the battery (CR2032) within 20 readings or one week.

Refer to your user guide for complete troubleshooting and error information, or call Bayer Customer Service:

1-800-348-8100
(24 hours/day)



Bayer HealthCare

Bayer HealthCare LLC
Mishawaka, IN 46544 USA
www.bayerdiabetes.com

©2006 Bayer
50184413 Rev. 2/08



Breeze²

BLOOD GLUCOSE MONITORING SYSTEM



QUICK REFERENCE GUIDE