



Mononucleosis Quiz

Circle the correct answer

This quiz is an educational tool intended to assist facilities in evaluating their operators' understanding of the QuickVue+ Infectious Mononucleosis Test procedure. This quiz is not intended to be used as sole evidence of operator training or competency. Facilities are responsible for ensuring the quality of the testing performed by their operators. When testing controls or patient specimens, follow the current Package Insert instructions and/or Procedure Card included in the kit.

1. How often does the manufacturer recommend that external controls be run on the kit?

- a) Once per kit
- b) Once per each new shipment of kits
- c) Every first sample tested each day
- d) Once per each untrained operator, once for each new shipment of kits – provided that each different lot in a shipment is tested

2. At what temperature do the kits need to be stored?

- a) 2°C to 8°C (refrigerated)
- b) 15°C to 30°C (room temperature)
- c) 2°C to 30°C (refrigerated or room temperature)
- d) -20°C (frozen)

3. What sample types may be used with this kit?

- a) Whole blood
- b) Serum
- c) Plasma
- d) All of the above

4. What is the read time for this kit?

- a) 5 minutes
- b) 10 minutes
- c) When ANY blue color appears in the Test Complete Window (approximately 5 minutes)
- d) 20 minutes

5. After how long should you NOT read test results?

- a) 5 minutes
- b) 10 minutes
- c) 15 minutes
- d) 20 minutes

6. How many drops of Developer do you add to the "Add" well after adding the patient sample?

- a) 10 drops
- b) 5 drops
- c) 8 drops
- d) 15 drops

7. If using fingertip blood, how many hanging drops should be added to the "Add" well?

- a) 1
- b) 2
- c) 3
- d) 4

8. When should the Test Cassette be removed from the foil pouch?

- a) 5 minutes before use
- b) 1 hour before use
- c) 30 minutes before use
- d) Just before use

9. How long can serum and plasma samples be stored before testing?

- a) 24 hours at room temperature
- b) 72 hours at 2°C to 8°C
- c) Below -20°C for 3 months
- d) Both b and c

10. Can commercial external controls other than Quidel's be used with this kit?

- a) Yes
- b) No

Mononucleosis Quiz

Circle the correct answer

This quiz is an educational tool intended to assist facilities in evaluating their operators' understanding of the QuickVue+ Infectious Mononucleosis Test procedure. This quiz is not intended to be used as sole evidence of operator training or competency. Facilities are responsible for ensuring the quality of the testing performed by their operators. When testing controls or patient specimens, follow the current Package Insert instructions and/or Procedure Card included in the kit.

1. How often does the manufacturer recommend that external controls be run on the kit?

- a) Once per kit
- b) Once per each new shipment of kits
- c) Every first sample tested each day
- d) Once per each untrained operator, once for each new shipment of kits – provided that each different lot in a shipment is tested

2. At what temperature do the kits need to be stored?

- a) 2°C to 8°C (refrigerated)
- b) 15°C to 30°C (room temperature)
- c) 2°C to 30°C (refrigerated or room temperature)
- d) -20°C (frozen)

3. What sample types may be used with this kit?

- a) Whole blood
- b) Serum
- c) Plasma
- d) All of the above

4. What is the read time for this kit?

- a) 5 minutes
- b) 10 minutes
- c) When ANY blue color appears in the Test Complete Window (approximately 5 minutes)
- d) 20 minutes

5. After how long should you NOT read test results?

- a) 5 minutes
- b) 10 minutes
- c) 15 minutes
- d) 20 minutes

6. How many drops of Developer do you add to the "Add" well after adding the patient sample?

- a) 10 drops
- b) 5 drops
- c) 8 drops
- d) 15 drops

7. If using fingertip blood, how many hanging drops should be added to the "Add" well?

- a) 1
- b) 2
- c) 3
- d) 4

8. When should the Test Cassette be removed from the foil pouch?

- a) 5 minutes before use
- b) 1 hour before use
- c) 30 minutes before use
- d) Just before use

9. How long can serum and plasma samples be stored before testing?

- a) 24 hours at room temperature
- b) 72 hours at 2°C to 8°C
- c) Below -20°C for 3 months
- d) Both b and c

10. Can commercial external controls other than Quidel's be used with this kit?

- a) Yes
- b) No