

MICROLUBE™ Instrument Lubricant



MicroLube™ is specially formulated to be used as a lubricant and rust inhibitor for all types of surgical and dental instruments including carbon steel. It is compatible with all sterilization methods and if stored properly, can be reused for 28 days without becoming contaminated.

- **Reduces likelihood of rust, tarnish or corrosion.**
- **Extends life of instruments, resulting in fewer repairs.**
- **Patient-safe lubricant will not cause abscesses in body.**

Instrument Lubricant

Order No.	Size	Physical Properties
T25	5 Gallons (19 Liters)	Form: Liquid Color: White Odor: None pH Concentration: 6.5

Uses	Temperature
<input type="checkbox"/> Presoak <input checked="" type="checkbox"/> Automatic Washer <input checked="" type="checkbox"/> Manual <input type="checkbox"/> Automatic Endoscope Reprocessor <input type="checkbox"/> Ultrasonic <input type="checkbox"/> Cart Washer	Usage: Room Temperature Storage: Cool, dry area

Product Features	Consumer Benefits
Rust Inhibitors:	Reduces likelihood of rust, tarnish or corrosion. Extends life of instruments, resulting in fewer repairs.
Neutral pH:	Product safe and compatible with all metals.
Non-silicone Lubricant:	Patient-safe lubricant will not cause abscesses in body.
Bacteriostatic:	If stored properly, not favorable to bacterial growth.
Environmentally-Friendly:	Phosphate-free formula.

Procedures	Directions for Use
Directions:	Wear appropriate personal protective equipment based on safe work practices.
Manual Lubrication:	Add 1 part lubricant to 6 parts distilled or demineralized water. If mixed lubricant is covered between use, solution can be changed every 28 days; however, if lubricant is used frequently, it is recommended to change the mixture each week. Use instrument lubricant daily to obtain maximum results.
Automatic Washing Equipment:	Add 1/8 fl. oz. (3.7 mL) to 2 fl. oz. (59 mL) per gallon (3.8 L) of water. Dilutions vary depending on temperature, water quality and equipment.