




Clarity BG1000

Blood Glucose Monitoring System Quick Reference Manual



Monitoring Blood Glucose




Insert the three black lines at the end of the test strip into the meter's strip slot until meter beeps.




Make sure the code on the screen matches the code number on the test strip vial. If it does not, then use another test strip. If the problem constantly occurs, please do not use test strips and contact Clarity Diagnostics Customer Service at 1-877-722-6339. www.claritydiagnostics.com

3




When the flashing blood symbol appears, the meter is ready for you to apply blood to the test strip.

4



Use your lancing device to obtain a blood drop.



Touch the blood drop to the front edge of the test strip sample channel. The blood is drawn into the test strip automatically. Continue to touch the blood drop at the edge of the test strip sample channel until you hear a beep sound.



After the beep sound, your blood test result will appear in 5 seconds. Discard the used test strip.

Reviewing the Results


Begin with the meter off condition. Press the 'M' button to turn the meter on.




Your last 14 days average result will be displayed. Press the 'C' button then the meter will display your most recent result. If you press 'M' button then the meter will turn off.




Screen Messages and Troubleshooting




Test strip is used or damaged.
Replace the new test strip.



Blood sample was applied to the test strip before the blood symbol () appeared on the Screen.



Temperature is too hot or too cold for the system to work properly.



The Blood sample is not filled enough for measurement because of abnormally high viscosity or insufficient volume.



The image shows a digital display with the error code 'Er5' in large black digits. Below the error code, the time '12:28' and '06:28 PM' is displayed in a smaller font.

12:28 06:28 PM

You put the wrong strip.

Please check if you are using the proper strip for the meter.



The image shows a digital display with the error code 'Er6' in large black digits. Below the error code, the time '12:28' and '06:28 PM' is displayed in a smaller font.

12:28 06:28 PM

This error appears when meter itself finds out internal problems.

Please contact Clarity Diagnostics Customer Service at 1-877-722-6339.



The image shows a digital display with the 'HI' mark in large black digits. Below the mark, the unit 'mg/dL' is displayed in a smaller font. At the bottom, the time '12:28' and '06:28 PM' is displayed.

mg/dL

12:28 06:28 PM

When the test result is higher than 600mg/dL, the screen will show the 'Hi' mark. It indicates that the value of your blood glucose is too high.



The image shows a digital display with the 'Lo' mark in large black digits. Below the mark, the unit 'mg/dL' is displayed in a smaller font. At the bottom, the time '12:28' and '06:28 PM' is displayed.

mg/dL

12:28 06:28 PM

When the test result is lower than 20mg/dL, the screen will show the 'Lo' mark. It indicates that the value of your blood glucose is too low.

Clarity Diagnostics, LLC, Boca Raton, FL USA 33487
Customer Service: 1-877-722-6339 www.claritydiagnostics.com

© Clarity Diagnostics, LLC. All rights reserved. Manufactured for Clarity Diagnostics, LLC.
Made in China

REV06102016
XP1044370A Rev.12.07.16